



CERTIFICATE NUMBER
18-HS1779163-PDA

DATE
26 Sep 2018

ABS TECHNICAL OFFICE
Houston ESD - Piping

CERTIFICATE OF Design Assessment

This is to certify that a representative of this Bureau did, at the request of
RKI INSTRUMENTS, INC.

assess design plans and data for the below listed product. This assessment is a representation by the Bureau as to the degree of compliance the design exhibits with applicable sections of the Rules. This assessment does not waive unit certification or classification procedures required by ABS Rules for products to be installed in ABS classed vessels or facilities. This certificate, by itself, does not reflect that the product is Type Approved. The scope and limitations of this assessment are detailed on the pages attached to this certificate.

Product: **Gas Monitor**
Model: **BEACON 200**

This Product Design Assessment (PDA) Certificate 18-HS1779163-PDA, dated 26/Sep/2018 remains valid until 25/Sep/2023 or until the Rules or specifications used in the assessment are revised (whichever occurs first).

This PDA is intended for a product to be installed on an ABS classed vessel, MODU or facility which is in existence or under contract for construction on the date of the ABS Rules or specifications used to evaluate the Product.

Use of the Product on an ABS classed vessel, MODU or facility which is contracted after the validity date of the ABS Rules and specifications used to evaluate the Product, will require re-evaluation of the PDA.

Use of the Product for non ABS classed vessels, MODUs or facilities is to be to an agreement between the manufacturer and intended client.

AMERICAN BUREAU OF SHIPPING

Tim Kimble
Engineer/Consultant

NOTE: This certificate evidences compliance with one or more of the Rules, Codes, Standards or other criteria of ABS or a statutory, industrial or manufacturer's standards. It is issued solely for the use of ABS, its committees, its clients or other authorized entities. Any significant changes to the above mentioned product without approval from ABS will result in this certificate becoming null and void. This certificate is governed by the terms and conditions as contained in ABS Rules 1-1-A3, 5.9 Terms and Conditions of the Request for Product Type Approval and Agreement (2016).

RKI INSTRUMENTS

33248 CENTRAL AVE.

UNION CITY CA

United States 94587

Telephone: 800-754-5165

Fax: 510-441-5650

Email: sales@rkiinstruments.com

Web: rkiinstruments.com

Tier: 3 - Type Approved, unit certification not required

Product: Gas Monitor
Model: BEACON 200
Intended Service:
Marine & Offshore Application.

Description:
Two Channel Wall Mount Controller.

Rating:
Power: 115/220 VAC or 24 VDC.
Operating Temperature: -4°F to 122°F (-20°C to 50°C).
Storage Temperature: -4°F to 158°F (-20°C to 70°C).
Enclosure: NEMA 4X.
See attached "Beacon200, Beacon 200 Product Brochure - F018-1202-2000" Data Sheet for details.

Service Restriction:
- Unit Certification is not required for this product. If the manufacturer or purchaser request an ABS Certificate for compliance with a specification or standard, the specification or standard, including inspection standards and tolerances, must be clearly defined.
- Suitable for installed in non-hazardous areas only.

Comments:
- The Manufacturer has provided a declaration about the control of, or the lack of Asbestos in this product.
- Sensors wired remote from the Controller and installed in a Hazardous Area are to be Explosion Proof and detailed installation drawings are to be submitted for review.

Notes/Drawing/Documentation:
Document No. RKI Instruments Project 70048823, CSA Certificate of Compliance No. 70048823, Dated 04 November 2015, Pages: 32
Document No. beacon200, Beacon 200 Product Brochure - F018-1202-2000, Pages: 2

Terms of Validity:
This Product Design Assessment (PDA) Certificate 18-HS1779163-PDA, dated 26/Sep/2018 remains valid until 25/Sep/2023 or until the Rules or specifications used in the assessment are revised (whichever occurs first).

This PDA is intended for a product to be installed on an ABS classed vessel, MODU or facility which is in existence or under contract for construction on the date of the ABS Rules or specifications used to evaluate the Product.

Use of the Product on an ABS classed vessel, MODU or facility which is contracted after the validity date of the ABS Rules and specifications used to evaluate the Product, will require re-evaluation of the PDA.

Use of the Product for non ABS classed vessels, MODUs or facilities is to be to an agreement between the manufacturer and intended client.

STANDARDS

ABS Rules:
Rules for Conditions of Classification, Part 1- 2018 Steel Vessels Rules 1-1-4/7.7, 1-1-A3, 1-1-A4, which covers the following:
2018 Steel Vessels Rules: 4-8-3/1.11.1, 4-8-3/Table 2

RKI INSTRUMENTS
33248 CENTRAL AVE.
UNION CITY CA
United States 94587
Telephone: 800-754-5165
Fax: 510-441-5650
Email: sales@rkiinstruments.com
Web: rkiinstruments.com

Tier: 3 - Type Approved, unit certification not required

Rules for Conditions of Classification, Part 1- 2018 Offshore Units and Structures 1-1-4/9.7, 1-1-A2, 1-1-A3, which covers the following:

2018 Mobile Offshore Drilling Units Rules: 4-3-3/3.1.1(a), 4-3-3/Table 1

National:

CSA C22.2 No. 61010-1-12
ANSI/UL Std. No. 61010-1 (3rd Edition)

International:

NA

Government:

NA

EUMED:

NA

OTHERS:

NA